



BID BULLETIN No. 1

This **Bid Bulletin No. 1** is hereby issued for the information and guidance of all participating bidders for the **Procurement of Real-Time PCR System Compatible with Ion Gene Studio S5 Prime System– Ref. # 2025-05-11**

The following changes shall take effect based on the discussions during pre-bid conference:

Specifications:

- excitation/detection range for 96 wells: 450-680nm/500-730nm **or better**

Inclusions:

- 1 unit 2 KVA UPS with built-in AVR **or separate units**
- One day on-site applications training by a **certified** Field Applications Specialist

This Bid Bulletin No.1 is issued to modify or amend items in the bid documents.

For guidance and information of all concerned.

ALEXIS M. ORTIZ

Chairperson

DOST-FNRI Bids and Awards Committee